

CITY OF AUSTIN
PURCHASE SPECIFICATIONS
FOR
Request for Quote MHJ0016
Breast Pumps and Kits

1.0 SCOPE AND CLASSIFICATION

This specification describes and establishes the minimum standard requirements for Ameda Elite and Purely Yours Breast Pumps and Sterile Hygiene Kit with Flange Inserts.

2.0 APPLICABLE SPECIFICATIONS

2.1. Ameda Purely Yours Breast Pump, Dual Hygienikit Milk Collection System

- 2.1.1. Shall be dual hygienikit milk collection system
- 2.1.2. Shall include a microfiber tote bag, two 30.5 millimeter breast flanges, two 28.5 millimeter reducing inserts, two 22.5 millimeter inserts and two 25 millimeter flanges, AC power adapter, cool and carry tote with 3 cooling elements, six 4 oz bottles with lids, and one hand handle assembly to convert hygienikit to manual pump, milk storage guidelines card, and instructional DVD.
- 2.1.3. Shall be BPA and DEHP free.

2.2 Ameda Elite Breast Pump

- 2.2.1. Shall be hospital grade.
- 2.2.2. Shall be independently adjustable suction and speed settings.
- 2.2.3. Shall have built in bottle and flange holders.
- 2.2.4. Shall be able to be mounted on Ameda trolley.
- 2.2.5. Shall include breast pump carry bag or storage container.
- 2.2.6. Shall have at least 1 year warranty for refurbished and 2 year warranty for new.
- 2.2.7. Shall be BPA and DEHP free.

2.3 Sterile Hygiene Kit with Flange Inserts

- 2.3.1. Shall include one tubing adaptor, two silicone tubings, two adaptor caps, two silicone diaphragms, two 30.5 millimeter breast flanges, two 28.5 millimeter reducing inserts, two 22.5 millimeter inserts and two 25 millimeter flanges
- 2.3.2. Shall be BPA and DEHP free.
- 2.3.3. Shall be sterile.
- 2.3.4. Shall be compatible with Ameda breast pumps.